



TiterZyme® EIA

human IFN- γ

Enzyme Immunometric Assay Kit

Catalog No. 900-136

96 Well Kit

Table of Contents

Description	Page	2
Introduction		2
Precautions		2
Materials Supplied		3
Storage		3
Materials Needed but Not Supplied		3
Sample Handling		4
Procedural Notes		5
Reagent Preparation		5
Assay Procedure		6
Calculation of Results		7
Typical Results		7
Typical Standard Curve		8
Calibration		8
Performance Characteristics		9
Sample Dilution Recommendations		11
References		11
Limited Warranty		12

FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

Description

Assay Designs' human IFN- γ TiterZyme[®] Enzyme Immunometric Assay (EIA) kit is a complete kit for the quantitative determination of human IFN- γ in biological fluids. Please read the complete kit insert before performing this assay. The kit uses an antibody to human IFN- γ immobilized on a microtiter plate to bind the human IFN- γ in the standards or sample. A recombinant human IFN- γ Standard is provided in the kit. A biotinylated antibody is added to the standards and samples. This labeled antibody binds to the human IFN- γ captured on the plate. After a short incubation the excess standard, sample and antibody are washed out and streptavidin conjugated to Horseradish peroxidase is added, which binds to the biotinylated human IFN- γ antibody. Excess conjugate is washed out and substrate is added. The substrate reacts with the conjugate bound to the IFN- γ captured on the plate. After a short incubation, the enzyme reaction is stopped and the color generated is read at 450 nm. The measured optical density is directly proportional to the concentration of human IFN- γ in either standards or samples. For further explanation of the principles and practices of immunoassays please see the excellent books by Chard¹ or Tijssen².

Introduction

Human Interferon-gamma (IFN- γ) is a polypeptide of 166 amino acids including the signal sequence of 23 amino acids. Two active forms of the protein have been found with different glycosylation sites. The first form is glycosylated only at position 25, while the second is glycosylated at positions 25 and 97.⁴⁻⁸ IFN- γ is alternatively known as antigen-induced interferon, immune interferon (IIF), Type-2 interferon, T interferon, mitogen-induced interferon, and pH2-labile interferon.⁸ IFN- γ plays an important role in immune response. It is secreted by Th1 cells which in turn activates macrophage through the CD40 surface ligand. This secretion also inhibits the development of Th2 cells.³ Human and mouse IFN- γ share only a 40% homology on the amino acid level.⁹

Precautions

FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

1. Stop Solution is a 0.18 M sulfuric acid solution. This solution is caustic; care should be taken in use.
2. The activity of the Horseradish peroxidase conjugate is affected by nucleophiles such as azide, cyanide and hydroxylamine.
3. We test this kit's performance with a variety of samples, however it is possible that high levels of interfering substances may cause variation in assay results.
4. The human IFN- γ Standard provided, Catalog No. 80-1289, should be handled with care because of the known and unknown effects of IFN- γ .

Materials Supplied

- 1. human IFN- γ Microtiter Plate, One Plate of 96 Wells, Catalog No. 80-1288**
A strip microtiter plate coated with antibody specific to human IFN- γ .
- 2. human IFN- γ Antibody , 8 mL, Catalog No. 80-1291**
A solution of biotinylated antibody to human IFN- γ .
- 3. human IFN- γ Standard Diluent, 12 mL, Catalog No. 80-1290**
Phosphate buffered saline containing proteins and antibiotics.
- 4. human IFN- γ Streptavidin-HRP Concentrate, 75 μ L, Catalog No. 80-1292**
A concentrated solution of streptavidin conjugated to Horseradish peroxidase.
- 5. human IFN- γ Streptavidin-HRP Dilution Buffer, 14 mL, Catalog No. 80-1293**
- 6. Wash Buffer Concentrate, 50 mL, Catalog No. 80-1253**
Tris buffered saline containing detergents.
- 7. human IFN- γ Standard, 2 vials, Catalog No. 80-1289**
Two vials of lyophilized recombinant human IFN- γ .
- 8. human IFN- γ TMB Substrate, 13 mL, Catalog No. 80-1294**
A solution of 3,3',5,5' tetramethylbenzidine (TMB) and hydrogen peroxide. **Protect from prolonged exposure to light.**
- 9. human IFN- γ Stop Solution, 14 mL, Catalog No. 80-1295**
A 0.18 M solution of sulfuric acid in water. Keep tightly capped. Caution: **Caustic.**
- 10. human IFN- γ Assay Layout Sheet, 1 each, Catalog No. 30-0212**
- 11. Plate Sealer, 3 each, Catalog No. 30-0012**

Storage

All components of this kit are stable at 4 °C until the kit's expiration date.

Materials Needed but Not Supplied

1. Deionized or distilled water. No difference in assay results is seen with distilled water.
2. Precision pipets for volumes between 50 μ L and 1,000 μ L.
3. Repeater pipet for dispensing 50 and 100 μ L.
4. Disposable beakers for diluting buffer concentrates.
5. Graduated cylinders.
6. Adsorbent paper for blotting.
7. Microcentrifuge to prepare Streptavidin-HRP Solution.
8. Microplate reader capable of reading at 450 nm, preferably with correction between 570 nm and 590 nm.
9. Log-log graph paper for plotting the standard curve.

Sample Handling

Assay Designs' TiterZyme[®] Enzyme Immunometric Assay is compatible with human IFN- γ samples in a wide range of matrices.

Culture fluids, serum, EDTA, heparin and sodium citrate plasma and urine are suitable for use in the assay. Samples containing a visible precipitate must be clarified prior to use in the assay. Do not use grossly hemolyzed or lipemic specimens. Samples in the majority of tissue culture media can also be read in the assay, provided the standards have been diluted into the tissue culture media instead of Standard Diluent. There will be a small change in binding associated with running the standards and samples in media. Users should only use standard curves generated in media, or Standard Diluent to calculate concentrations of human IFN- γ in the appropriate matrix.

Samples must be stored frozen to avoid loss of bioactive human IFN- γ . If samples are to be run within 24 hours, they may be stored at 4 °C. Otherwise, samples must be stored frozen at or below -70 °C to avoid loss of bioactive human IFN- γ . Up to three freeze/thaw cycles of serum has been shown to have no effect on human IFN- γ levels. Nonetheless, excessive freeze/thaw cycles should be avoided. Prior to assay, frozen sera should be brought to room temperature slowly and gently mixed by hand. Do not thaw samples in a 37 °C incubator. Do not vortex or sharply agitate samples.

Procedural Notes

1. Do not mix components from different kit lots or use reagents beyond the kit expiration date.
2. Allow all reagents to warm to room temperature for at least 30 minutes before opening.
3. Standards must be prepared in polypropylene or polyethylene tubes. Do not use polystyrene, polycarbonate or glass tubes.
4. Pre-rinse the pipet tip with reagent, use fresh pipet tips for each sample, standard and reagent.
5. Pipet standards and samples to the bottom of the wells.
6. Add the reagents to the side of the well to avoid contamination.
7. This kit uses plates with removable strips. Unused strips must be kept desiccated at 4 °C in the sealed bag provided. The wells should be used in the frame provided.
8. **Prior to addition of conjugate and substrate, ensure that there is no residual wash buffer in the wells. Any remaining wash buffer may cause variation in assay results.**
9. **It is important that the matrix for the standards and samples be as similar as possible. Human IFN- γ samples diluted with Standard Diluent should be run with a standard curve diluted in the same buffer. Serum, plasma and urine samples should be evaluated against a standard curve run in Standard Diluent while Tissue Culture samples should be read against a standard curve diluted in the same complete but non-conditioned media. See Reagent Preparation, step #2.**

Reagent Preparation

1. Wash Buffer

Prepare the Wash Buffer by diluting 50 mL of the supplied concentrate with 1,450 mL of deionized water. This can be stored at room temperature until the kit expiration, or for 3 months, whichever is earlier.

2. human IFN- γ Standards

Reconstitute standard with deionized water. Reconstitution volume is stated on the standard vial label. Let it sit at room temperature for 5 minutes. Mix gently. This solution contains 2,500 pg/mL human IFN- γ . When testing serum, plasma or urine samples, use the Standard Diluent provided to prepare standard curve serial dilutions. When using cell culture supernatants use tissue culture media to prepare the standard curve serial dilutions.

Label five 12x75 mm test tubes #1 through #5. Pipet 240 μ L of Standard Diluent or tissue culture media into tubes #1 through #5. Add 160 μ L of the 2,500 pg/mL Standard to tube #1. Vortex thoroughly. Add 160 μ L of tube #1 to tube #2 and vortex thoroughly. Add 160 μ L of tube #2 to #3 and vortex thoroughly. Continue this for tubes #4 and #5.

The concentration of human IFN- γ in tubes #1 through #5 will be 1,000, 400, 160, 64 and 25.6 pg/mL respectively. See human IFN γ Assay Layout Sheet for dilution details.

Diluted standards should be used within 60 minutes of preparation. Do not store reconstituted standards.

3. Streptavidin-HRP Solution

Prepare Streptavidin-HRP solution **immediately before use**. Do not store prepared Streptavidin-HRP solution. Use a plastic tube to prepare Streptavidin-HRP solution. Briefly centrifuge the Streptavidin-HRP Concentrate to force entire vial contents to the bottom. For each strip used, mix 2.5 μL of Streptavidin-HRP Concentrate with 1 mL of Streptavidin-HRP Dilution Buffer.

Assay Procedure

Bring all reagents to room temperature for at least 30 minutes prior to opening.

All standards, controls and samples should be run in duplicate.

1. Refer to the Assay Layout Sheet to determine the number of strips to be used and put any remaining strips with the desiccant back into the pouch and seal the ziploc. Store unused wells at 4 °C.
2. Pipet 50 μL of the Biotinylated Antibody into each well, except the Blank.
3. Pipet 50 μL of standard diluent or Tissue Culture Media into the S0 (0 pg/mL standard) wells.
4. Pipet 50 μL of Standards #1 through #5 into the appropriate wells.
5. Pipet 50 μL of the Samples into the appropriate wells.
6. Tap the plate gently to mix the contents, and seal with the plate sealer provided.
7. Incubate the plate at room temperature for 2 hours.
8. Empty the contents of the wells and wash by adding 400 μL of wash solution to every well. Repeat the wash 2 more times for a total of **3 washes**. After the final wash, empty or aspirate the wells and firmly tap the plate on a lint free paper towel to remove any remaining wash buffer.
9. Add 100 μL of the freshly prepared Streptavidin-HRP Conjugate to each well, except the Blank.
10. Seal the plate and incubate at room temperature for 30 minutes.
11. Empty the contents of the wells and wash by adding 400 μL of wash solution to every well. Repeat the wash 2 more times for a total of **3 washes**. After the final wash, empty or aspirate the wells and firmly tap the plate on a lint free paper towel to remove any remaining wash buffer.
12. Pipet 100 μL of Substrate Solution into each well.
13. Incubate for 30 minutes at room temperature in the dark.
14. Pipet 100 μL Stop Solution to each well.
15. Blank the plate reader against the Blank wells, read the optical density at 450 nm, preferably with correction between 570 and 590 nm. If the plate reader is not able to be blanked against the Blank wells, manually subtract the mean optical density of the Blank wells from all the readings.

Calculation of Results

Several options are available for the calculation of the concentration of human IFN- γ in the samples. We recommend that the data be handled by an immunoassay software package utilizing a 4 parameter logistic curve fitting program. If data reduction software is not readily available, the concentration of human IFN- γ can be calculated as follows:

1. Calculate the average net Optical Density (OD) bound for each standard and sample by subtracting the average Blank OD from the average OD for each standard and sample.

$$\text{Average Net OD} = \text{Average OD} - \text{Average Blank OD}$$

2. Using linear graph paper, plot the Average Net OD for each standard versus concentration of human IFN- γ in each standard. Approximate a straight line through the points. The concentration of h IFN- γ in the unknowns can be determined by interpolation.


Typical Results

The results shown below are for illustration only and **should not** be used to calculate results from another assay.

<u>Sample</u>	<u>Average OD</u>	<u>Net OD</u>	h IFN-γ (pg/mL)
Blank	(0.043)		
S0	0.097	0.054	0
S1	1.742	1.699	1,000
S2	0.788	0.745	400
S3	0.397	0.354	160
S4	0.221	0.178	64
S5	0.135	0.092	25.6
Unknown 1	1.405	1.362	783.7
Unknown 2	0.307	0.264	113.3

Typical Standard Curve

A typical standard curve is shown below. This curve **must not** be used to calculate human IFN- γ concentrations; each user must run a standard curve for each assay.



Calibration

The standards in this kit have been calibrated to the NIAID recombinant IFN-g standard lot Gxg23-902-535. One (1) pg of Assay Designs' Standard = 0.03 NIAID units.

Performance Characteristics

Sensitivity: < 2 pg/mL

The sensitivity or Lower Limit of Detection (LLD) is determined by assaying replicates of zero and the standard curve. The mean signal of zero + 2 standard deviations read in dose from the standard curve is the LLD. This value is the smallest dose that is not zero with 95% confidence.

Linearity

A sample containing 686.9 pg/mL human IFN- γ was serially diluted 4 times 1:2 in the standard diluent supplied in the kit and measured in the assay. The data was plotted graphically as actual human IFN- γ concentration versus measured human IFN- γ concentration.

The line obtained had a slope of 1.046 with a correlation coefficient of 0.998.

Precision

Intra-assay CV: < 10%

Inter-assay CV: < 10%

Cross Reactivities

The TiterZyme[®] human IFN- γ EIA Kit is specific for natural and recombinant human IFN- γ . It is unaffected by the presence of human IL-1 α , IL-1 β , IL-2, IL-3, IL-6, IL-7, IL-8, TNF- α , or GM-CSF.

Sample Recoveries

Cytokine recovery is determined by spiking low and high levels of recombinant human IFN- γ in to normal human serum, plasma, and urine samples, and a Standard Diluent control buffer. Mean recoveries are as follows:

<u>Spike Level</u>	<u>900 pg/mL</u>	<u>150 pg/mL</u>
Mean Serum Recovery	106%	84%
Mean Plasma Recovery	95%	65%
Mean Urine Recovery	105%	80%

References

1. T. Chard, "An Introduction to Radioimmunoassay & Related Techniques 4th Ed.", (1990) Amsterdam: Elsevier.
2. P. Tijssen, "Practice & Theory of Enzyme Immunoassays", (1985) Amsterdam: Elsevier.
3. Alberts, et al., "The Cell", (2002) New York, Garland Science: 1412-1415.
4. P. Chomarat, et al., "The Cytokine Handbook", (1998) Academic Press: 491-508.
5. E Rinderknecht, et al., J. Biol. Chem., (1984) 259(11): 6790-7.
6. R Devos, et al., Nucleic Acids Res., (1982) 10(8): 2487-501.
7. YK Yip, et al., Proc. Natl. Acad. Sci. USA, (1982) 79(6): 1820-4.
8. Cytokines Online Pathfinder Encyclopaedia, "IFN-gamma".
9. <http://www.copewithcytokines.de/cope.cgi?3172> Accessed 23 July, 2002.
10. P.W. Gray and D.V. Goeddel, Proc. Natl. Acad. Sci. USA, (1983) 80(19): 5842-6.

LIMITED WARRANTY

Assay Designs, Inc. warrants that at the time of shipment this product is free from defects in materials and workmanship. This warranty is in lieu of any other warranty expressed or implied, including but not limited to, any implied warranty of merchantability or fitness for a particular purpose.

Assay Designs must be notified of any breach of this warranty within 48 hours of receipt of the product. No claim shall be honored if Assay Designs is not notified within this time period, or if the product has been stored in any way other than outlined in this publication. The sole and exclusive remedy of the customer for any liability based upon this warranty is limited to the replacement of the product, or refund of the invoice price of the goods.



For more details concerning the information within this kit insert, or to order any of Assay Designs' products, please call (734) 668-6113 between 8:30 a.m. and 5:30 p.m. EST. Orders or technical questions can also be transmitted by fax or e-mail 24 hours a day.

Material Safety Data Sheet (MSDS) available on our website or by fax.

**Assay Designs, Inc.
800 Technology Drive
Ann Arbor, MI 48108
U.S.A.**

**Telephone: (734) 668-6113
(800) 833-8651 (USA & Canada only)
Fax: (734) 668-2793
e-mail: info@assaydesigns.com
website: www.assaydesigns.com**

Simplify Your Science®

Catalog No. 25-0544

© 2004



May 3, 2005

